


HPLC DATA SHEET

Olopatadine Hydrochloride (USP, Assay)



1. Olopatadine

HPLC condition

Column	Aegispak C8 (5um, 4.6mm I.d. × 150mm)
Mobile Phase	Buffer : Acetonitrile (28 : 72) * Buffer : Dissolve 13.6g of monobasic potassium phosphate in 1L of water, add 1mL of triethylamine, and mix. Adjust with phosphoric to a pH 3.0
Flow rate	1.0mL/min
Temperature	-
Detector	UV (299nm)