


# HPLC DATA SHEET

Olopatadine Hydrochloride (USP, Limit of Late Eluting Impurities)



1. Olopatadine
2. Olopatadine related compound C

## HPLC condition

Column Aegispak C8 (5 $\mu$ m, 4.6mm I.d.  $\times$  150mm)

Buffer : Acetonitrile (1 : 1)

Mobile Phase \* Buffer : Dissolve 13.6g of monobasic potassium phosphate in 1L of water, add 1mL of triethylamine, and mix. Adjust with phosphoric to a pH 3.0

Flow rate 1.0mL/min

Temperature -

Detector UV (299nm)